

# Atos

## DECLARATION OF CONFORMITY FOR THE PRODUCT(S)

### Freevent Dressing Foam

**Basic UDI-DI: 7331791-COM-0-000-0004-5B**

This Declaration of Conformity is issued under the sole responsibility of Atos Medical AB. Atos Medical AB declares that the devices covered by the present declaration are in conformity with the Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices and the requirements specified herein are fulfilled and, clause 6.6 of Schedule 3 to the Australian Therapeutic Goods (Medical Devices) Regulations 2002 including amendments in effect at the issuance date.

**Intended use/purpose:**

Freevent Dressing Foams are single use tracheostomy dressings that provide protection between the tracheal cannula and the skin and absorb secretions.

Hörby, Sweden, date as stated on last page



.....  
Martin Richardson, Senior Vice President Operations & Quality  
on behalf of the CEO of Atos Medical AB.

**Manufacturer:** Atos Medical AB  
Kraftgatan 8, SE-242 35 Hörby  
Sweden

Telephone: +46 (0)415 198 00  
Email: Info@atosmedical.com  
Web: www.atosmedical.com

**SRN number:** SE-MF-00000725

**Competent Authority** Medical Products Agency  
Sweden

# DECLARATION OF CONFORMITY

FOR THE PRODUCT(S)

**7331791-COM-0-000-0004-5B**

REF	Device name	Class*	GMDN code
1431	Freevent Dressing Foam 65x70	I	15624

\*Product Classification according to Annex VIII EU Medical Device Regulation 2017/745

#### **In compliance with Therapeutic Goods (Medical Devices) Regulations 2002**

Each kind of medical device to which the system has been applied complies with the applicable provisions of the essential principles, the classification rules, and the full quality assurance procedures, at each stage, from the design of the device until its final inspection before being supplied.

#### **In compliance with UK Medical Devices Regulations 2002 as amended**

Each kind of medical device to which the system has been applied fulfils the applicable obligations imposed by Annex VII and the device meets the provisions of the applicable section of the Medical Devices Regulations.

Authorized representative/UK Responsible Person:

Atos Medical UK Ltd Tottle Road  
Cartwright House  
Nottingham  
Nottinghamshire NG2 1RT  
England, United Kingdom

#### **Common Specification(s) as per Article 9, and other Union Legislation(s)**

- No relevant Common Specifications as defined in Article 9 to list.
- No relevant Union Legislations to list
- No European Representative

Document Approvals  
Approved Date: 2025-05-09

Approval Task Verdict: Approve	SEHRBHNU Ulrika Svensson, Regulatory Affairs Professional (ulrika.svensson-atosmedical@coloplast.com) Issuer 05-May-2025 09:17:56 GMT+0000
Approval Task Verdict: Approve	SEHRBPNH Håkan Persson, Quality Manager (hakan.persson-atosmedical@coloplast.com) Quality 07-May-2025 11:15:37 GMT+0000
Approval Task Verdict: Approve	MARRIC Martin Richardson, VP, International Manufacturing, Sweden, Germany & China (martin.richardson-atosmedical@coloplast.com) Management 09-May-2025 07:54:47 GMT+0000